

# Pressure Sensor PS3100 User Manual

Version 1.1



PS3100 User Manual



# Table of Contents

|          |   |           |
|----------|---|-----------|
| <b>1</b> | <b>Introduction.....</b>                        | <b>4</b>  |
| 1.1      | System Description .....                        | 4         |
| 1.2      | How to Use the Manual.....                      | 5         |
| 1.3      | Certification .....                             | 6         |
| 1.4      | Unpacking and Inspection.....                   | 6         |
| 1.5      | Serial Number .....                             | 6         |
| 1.6      | Warranty Policy .....                           | 7         |
| 1.7      | Factory Service & Repair .....                  | 7         |
| 1.8      | Contact Details.....                            | 8         |
| <b>2</b> | <b>Sensor Overview.....</b>                     | <b>9</b>  |
| 2.1      | Theory of Measurement.....                      | 9         |
| 2.1.1    | Ceramic Capacitive transducers.....             | 9         |
| 2.2      | Applications.....                               | 9         |
| 2.3      | Instrument Details .....                        | 10        |
| 2.3.1    | Sensor Design.....                              | 10        |
| 2.4      | Cable Details.....                              | 11        |
| 2.4.1    | Mechanical Specifications.....                  | 11        |
| 2.5      | Options and Accessories.....                    | 12        |
| 2.5.1    | Absolute or Gauge .....                         | 12        |
| 2.5.2    | Closed Venting System (CVS).....                | 12        |
| 2.5.3    | Protective Nose Cones .....                     | 13        |
| 2.5.4    | BSP Fittings .....                              | 13        |
| 2.5.5    | Cable Options (Inc Detachable Cables) .....     | 13        |
| 2.5.6    | Sensor Body.....                                | 13        |
| 2.5.7    | Communication Cables.....                       | 13        |
| 2.5.8    | Optional Serial Output – SDI Adapter Unit ..... | 14        |
| 2.5.9    | On Board Battery Housing.....                   | 14        |
| 2.6      | Battery Replacement.....                        | 15        |
| 2.6.1    | Important Battery Information .....             | 15        |
| 2.6.2    | Battery Warnings:.....                          | 15        |
| 2.7      | Sensor Factory Calibration.....                 | 16        |
| <b>3</b> | <b>Sensor Operation.....</b>                    | <b>17</b> |
| 3.1      | Wiring & Connections.....                       | 17        |
| 3.2      | Software Setup.....                             | 19        |
| 3.2.1    | Direct Link .....                               | 19        |
| 3.2.2    | SmartCom.....                                   | 19        |

|          |   |           |
|----------|---|-----------|
| 3.2.3    | Properties.....                                 | 24        |
| 3.2.4    | Channels.....                                   | 24        |
| 3.2.5    | External Battery: (Field Supply).....           | 24        |
| 3.2.6    | Internal Battery: (Battery Pack).....           | 25        |
| 3.2.7    | Battery Life & Memory Calculator.....           | 25        |
| 3.2.8    | Temperature.....                                | 26        |
| 3.2.9    | Pressure.....                                   | 26        |
| 3.2.10   | Schedules.....                                  | 26        |
| 3.2.11   | Add New Schedule (Wizard).....                  | 28        |
| 3.2.12   | Add New Schedule – Final Screen.....            | 29        |
| 3.2.13   | Averaging.....                                  | 30        |
| 3.2.14   | Data Variation (Event Based Logging).....       | 30        |
| 3.2.15   | User Calibration Screens.....                   | 31        |
| 3.2.16   | Setting a Datum.....                            | 31        |
| 3.2.17   | Clear Memory Tab.....                           | 32        |
| 3.2.18   | Set Clock.....                                  | 33        |
| 3.2.19   | PC Clock.....                                   | 33        |
| 3.2.20   | Manual Entry Clock.....                         | 33        |
| 3.2.21   | Start Logger.....                               | 34        |
| 3.2.22   | Collect Data.....                               | 35        |
| 3.2.23   | Aquagraph – Tabular View.....                   | 36        |
| 3.2.24   | Aquagraph – Graphical View.....                 | 36        |
| 3.2.25   | Aquagraph – Tools & Options.....                | 36        |
| 3.3      | Configuration & Testing.....                    | 37        |
| 3.3.1    | Quick Check and Re-Calibration Method.....      | 37        |
|          | Field Deployment Considerations.....            | 40        |
| 3.3.2    | Cabling Considerations.....                     | 40        |
| 3.3.3    | Typical Sensor Installations.....               | 41        |
| 3.3.4    | Field Installation must ensure:.....            | 41        |
| 3.3.5    | Other Considerations.....                       | 41        |
| 3.3.6    | Guidelines for cleaning equipment.....          | 41        |
| <b>4</b> | <b>Appendix A -Additional Information.....</b>  | <b>42</b> |
| 4.1      | Specifications.....                             | 42        |
| 4.2      | Quick Start Guide – PS3100 Pressure Sensor..... | 42        |
| 4.3      | Engineering Note – Detachable Cables.....       | 45        |

# I Introduction

## 1.1 System Description

Thank you for purchasing the Greenspan Pressure Sensor Model PS3100. This manual provides a guide to the configuration, operation and maintenance of the sensor to provide long term reliable and accurate monitoring.

The new **PS3100 Pressure Sensor** incorporates a ceramic capacitance transducer and represents the current state of the art depth/level sensor intended for long-term deployment with accurate data collection and on-board data storage.

The high over range pressure rating and long-term stability of the ceramic sensor make this product suitable for demanding pressure-monitoring applications.

The factory calibration utilises a 36-point procedure guaranteeing that the +/- 0.1%FS accuracy specification is met throughout the pressure range at any temperature within the operational range of the sensor.

Rapid data collection with a data interval selectable down to one second and 4 user defined schedules, which are able to start and stop each other, ensures maximum data collection during critical events. Subsequently, battery performance is maximised through minimal power consumption.

A large memory capacity (2 Mb) allows the PS3000 to store all data over long periods before being downloaded either locally or remotely by your preferred communications method. Memory wrap is user selectable ensuring that the most recent data is always available.

Output from the PS3000 is RS232 with an optional SDI-12 adapter.

The measurement units are user definable offering greater versatility to a wide range of applications.

Used in conjunction with a V-notch weir or flume, the PS3000 is able to monitor and record flow levels in numerous applications where other flow meters may not be suitable or are cost prohibitive.

All firmware upgrades can now be performed in the field without the need to return the sensor to the factory. User calibration is achievable via the SmartCom software package.

The Sensor is packaged in an Acetal or 316 Stainless Steel tube, with Double O ring connections for the cable and sensor head. This fully submersible sensor design is rugged and well proven and can withstand the harsh conditions found in remote field applications. The sensors are suitable for applications in harsh remote applications including groundwater, salty or acidic water conditions.

The PS3000 can be powered using external batteries (via the sensor cable) or with Greenspan's optional Lithium battery pack, which can be fitted as an extension to the sensor body, making it a fully self-contained measurement and logging system.

To avoid moisture ingress via the vent tube on gauge pressure sensors, a closed vent system is provided which maintains atmospheric pressure within the ceramic capacitance transducer while preventing moisture condensing within the sensor cable.

As an option, the PS3000 can be fitted with a detachable cable to allow ease of removal and shipping from the field.

## **1.2 How to Use the Manual**

Along with this manual, there are several other documents that may assist in the successful configuration and operation of the Greenspan PS3100 Sensor. These should be maintained on file as a permanent reference as to the features, applications and use of the PS3100.

***Greenspan PS3100 – Specifications Brochure***

***Greenspan PS3100 – Certificate of Conformance***

***Greenspan PS3100 – Quick Start Guide***

## 1.3 Certification

The PS3100 sensors are assembled and tested in accordance with Greenspan's ISO 9001 Quality Certified System. Each Sensor is individually manufactured and certified against a traceable Standard (refer Section 2.6)

Following calibration the sensors undergo a range of additional control processes to ensure that all specifications are consistent and documented.

- ***The instrument is visually inspected, marked and labelled.***
- ***The complete sensor calibration record is archived for reference, and batch number information is kept on file for statistical analysis.***
- ***An individual Certificate of Conformance is issued to the customer.***

## 1.4 Unpacking and Inspection

All Greenspan Analytical Sensors are made to order and are individually calibrated and inspected. This ensures that they leave the factory in a working condition. They are packed in new cartons for shipping. On receipt, the customer should inspect the packaging and contents for any signs of damage during transportation. The customer should also check that all items on the delivery note have been received.

Please contact the factory in case anything has been damaged or missing. A full set of documentation including Certificate of Conformance, Quick Start Guide, and Full Operator Manual will be provided with all equipment – either in hard copy format or in electronic format on the CD shipped with the goods.

If fitted with a 316 Stainless Steel body, the unit should only be used in low EC situations. Care should be taken against possible corrosion in high Chloride or Ferric solutions, water with high iron or sulphate reducing bacteria, or low dissolved oxygen. The Sensor can be fitted with an Acetal body which provides superior corrosion protection in a wide range of chemically active waters.

Because an individual sensor may be used in a variety of locations, media compatibility should be checked before installing and advice sought from Greenspan if any doubt exists.

## 1.5 Serial Number

Checking the Model Number and Range

Before installing your Greenspan PS3100 sensor check the information on the label is correct to confirm you have received the instrument you have ordered. The label will look similar to this.

|              |                 |
|--------------|-----------------|
| <b>MODEL</b> | <b>PS3100</b>   |
| <b>RANGE</b> | <b>0 – xx M</b> |
| <b>S/N</b>   | <b>012345</b>   |

The customer is advised to keep a record of the serial numbers in case the sensor is lost or the label damage. Greenspan Analytical keeps records of all sensors sold including a calibration history.

## 1.6 Warranty Policy

Greenspan Analytical warrants all new Greenspan products against defects in materials and workmanship for **12 months** from the date of invoice.

Products that prove to be defective during the warranty period will be repaired or replaced at the discretion of Greenspan Analytical.

Under Greenspan Analytical warranty conditions; it is the responsibility of the customer to cover shipping charges back to the factory. Upon repair/replacement Greenspan Analytical will cover the return shipping charges to the customer.

This warranty does not apply to products or parts thereof which have been altered or repaired outside of the Greenspan Analytical factory or other authorised service centre; or products damaged by improper installation or application, or subjected to misuse, abuse neglect or accident. This warranty also excludes items such as reference electrodes and Dissolved Oxygen membranes that may degrade during normal use.

Greenspan Analytical will not be liable for any incidental or consequential damage or expense incurred by the user due to partial or incomplete inoperability of it's products for any reason whatsoever or due to inaccurate information generated by its products.

All Warranty service will be completed as soon possible. If delays are unavoidable customers will be contacted immediately.

Any sensor should not be dismantled unless under instruction from Greenspan Analytical Technical Service staff. Incorrect handling will void the warranty.

## 1.7 Factory Service & Repair

The correct choice of sensor and assistance with field installation can be provided by Greenspan and their sales offices. A correct choice of equipment, together with technical advice and field experience should result in long term success in the field. **Greenspan Technical Services** is dedicated to customer support and provides assistance in the selection, installation, deployment and commissioning of sensors with a full range of consulting services. All Greenspan products are designed, developed and manufactured in Australia and can be supplied at short notice.

If for some reason sensors are required to be returned to our factory or your sales representative, please note the model and serial number, Describe the problem, including how and under what conditions the instrument was being used at the time of malfunction. Clean the product and cable. Decontaminate thoroughly if used in toxic or hazardous environment. Carefully pack product in original packaging if possible & include a statement certifying product and cable have been decontaminated with supporting information. Products returned for repair must be accompanied by a completed GRA (Goods Return Advice) form. All sensors returned for service and repair work must be properly decontaminated prior to return. A cleaning charge may be applied to sensors that require further decontamination. Service work will not commence until the quotation has been accepted by the customer. A purchase order for all repair and service work will be required before work is carried out.

## 1.8 Contact Details

### Australia

**Head Office**  
Goyen Controls Co Pty Ltd  
268 Milperra Road  
Milperra, NSW 2214

Telephone: 1800 805 372  
Facsimile: 1300 658 799

### Sales and Service

**Queensland, South Australia  
Victoria, Western Australia**

Telephone: 1800 805 372  
Facsimile: 1300 658 799

### USA

Goyen Valve Corporation  
1195 Airport Road  
Lakewood  
New Jersey 08701, USA

Telephone: 1 732 364 7800  
Facsimile: 1 732 364 1356

### Asia

Goyen Controls Co Pty Ltd  
Shanghai Representative Office  
1209 Greenland Business Centre  
1258 Yu Yuan Road  
Shanghai PC200050, CHINA

Telephone: 86 21 5239 8810  
Facsimile: 86 21 5239 8812

Goyen Controls Co Pty Ltd  
73-M Jalan Mega Mendung  
Kompleks Bandar OUG  
58200 Kuala Lumpur, MALAYSIA

Telephone: 60 37 987 6839  
Facsimile: 60 37 987 7839

Greenspan Singapore Pte Ltd  
02-01, Minwa Industrial Building  
39 Genting Lane  
Singapore 349554

Telephone: 65 6748 0140  
Facsimile: 65 6748 2534

### Europe

Goyen Controls Co UK Ltd  
Unit 3B Beechwood  
Chineham Business Park  
Basingstoke, Hampshire, RG24 8WA  
UNITED KINGDOM

Telephone: 44 1256 817 800  
Facsimile: 44 1256 843 164

**Tyco Umwelttechnik GmbH**  
Im Petersfeld 6  
D-65624 Altdenz  
GERMANY

Telephone: 49 6432 1001/1002  
Facsimile: 49 6432 63810

**Mecair S.r.l.**  
Via per Cinisello 97  
20054 Nova Milanese  
Milano,  
ITALY

Telephone: 39 362 375 118  
Facsimile: 39 362 375 124

### Address:

*Tyco Environmental Systems  
Greenspan Analytical Manufacturing Plant  
22 Palmerin Street  
WARWICK QLD 4370  
AUSTRALIA*

### Phone:

+ 61 (0)7 46601888

### Fax:

+ 61 (0)7 46601800

### Internet:

[www.tyco-environmental.com](http://www.tyco-environmental.com)

## 2 Sensor Overview

### 2.1 Theory of Measurement

The PS3000 Pressure Sensor utilizes a ceramic-based, capacitive element as the transducer. This is designed to be of rugged construction and incorporates active electronics as an integral part of the transducer substrate to enhance reliability and accuracy. Force applied to the ceramic element, due to the pressure, deforms its shape. This deformation causes a change in capacitance which can be measured by the electronics. The inherent stability and toughness of the ceramic ensures the repeatability and long term accuracy of the readings are maintained under the harshest field conditions.

The onboard microprocessor converts the transducer output voltage to a 16 bit digital signal and also measures the transducer temperature. This information is used to temperature compensate the sensor over the range 0 - 50°C. Both pressure and temperature are displayed in SmartCom in real units i.e. metres of depth and degrees centigrade.

#### 2.1.1 Ceramic Capacitive transducers

Benefits of the Ceramic Capacitance Sensors over other types of sensors are:

- Extremely high overload limit (typically up to 10 X overload protection)
- Absolute resistant to wear
- High temperature stability
- Excellent Long term stability
- Excellent Repeatability and linearity
- No hysteresis effects normally associated with Strain Type Sensors
- Corrosion resistant – Other sensors require contact of stainless steel face
- Not subject to mechanical fatigue that may affect strain gauge type sensors
- Low power consumption suitable for remote monitoring & control units

### 2.2 Applications

Applications in which the Greenspan PS3100 can be used include:


- Monitoring of streams and rivers.
- Monitoring of water storage bodies including stratification studies.
- Intermediate and final effluent treatment monitoring.
- Hydrological run off studies.
- Ground and bore water analysis.
- Drinking water filtration efficiency.
- Industrial process monitoring.
- Sludge and dredge monitoring.

## 2.3 Instrument Details

### 2.3.1 Sensor Design


The Greenspan PS3100 Sensor consists of the following primary elements:

- Ceramic capacitance transducer
- Microprocessor controlled signal conditioning and logging device
- Stainless steel or Acetal Body Material
- Data cable or Battery Pack option



## 2.4 Cable Details

All Greenspan Sensors utilise a specially designed Polyurethane Cable. The cable contains 12 x conductors, 1 x drain wire, and an internal vent tube. The outer jacket is made from UV stabilized Polyurethane and is suitable for all external, underwater or harsh environment applications. This common cable construction is utilized for vented and non vented sensors and all Greenspan Water Quality Sensors. Cables are generally factory fitted at time of manufacture in specified lengths. Cables can be joined or repaired in the field providing a waterproof connection can be maintained. Alternatively, cables can be terminated in waterproof junction boxes where cabling to other devices or longer cable runs are required.



### Cable Construction

- 1 – **Vent Tube:** Polyamide  
(size ID x OD) 2.40 x 3.20 mm
- 2 – **12 x Conductors**  
7 x 0.20 mm Tinned Copper  
Section = 0.22mm<sup>2</sup> AWG24  
Insulation: Polypropylene  
(size) = 1.10 mm ± 0.05 mm
- 3 – **Tape:** Polyester
- 4 – **Drain Wire:** 7x0.20 TinCu
- 5 – **Tape:** Polyester Aluminium
- 6 – **Jacket:** Polyurethane black,  
(size OD) 8.05 mm ± 0.15

### 2.4.1 Mechanical Specifications

- Specially Manufactured Greenspan Cable with 12 cores and Internal Vent
- High chemical resilience and abrasive resistance
- Conductor cross section : AWG 24,
- Electrical Resistance 9 ohm per 100m (per conductor)
- Operating temperature: 85°C (max.),
- Bending radius (static) : 6 ,
- Bending radius (dynamic) 12.
- Max Operating voltage : 250V
- Jacket Printing (white colour each meter)
- Conductor colour codes : green, yellow, white, black, brown, turquoise, violet, pink, red, blue, grey
- Tensile Strength is sufficient to self suspend the Greenspan Sensor to depths of 300m.
- Long term creep due to temperature effects or tensile loading is negligible.

The moulded cable is fitted to the sensor using a double o ring seal and located using 2 x grub screws. The length of the cable is not critical to the long term calibration and operation of the sensor (provided the electrical requirements such as minimum supply voltage are maintained).

## 2.5 Options and Accessories

### 2.5.1 Absolute or Gauge

Gauge Sensors are vented to atmosphere so that the effects of changes in barometric or atmospheric pressure do not affect water level readings. Sensors that are not vented to atmosphere are referred to as Absolute Sensors. The primary difference between the two types of sensors is the effect of atmospheric pressure on the water level measurements they provide.

Barometric Pressure acts on both sides of a Gauge sensor (ie via the water on one side and via the vent tube on the other). The Barometric pressure is cancelled out and has no effect on the water level readings. Gauge Sensors will read zero in air.

Barometric atmospheric pressure acts only on one side of a non vented or Absolute Sensor (on the water side). Any changes in Atmospheric pressure will be detected by the sensor and measured as changes in water pressure. As the Barometric pressure varies, these changes will be measured as water level changes even though the actual water level may have remained steady. Typical variations in Barometric Pressure when converted to head of water are in the order of +/- 100mm. A large change in Weather Pattern (Storm Front) may cause a drop in Barometric Pressure by up to 20Hpa which would cause an error of 200mm. Water level variations caused by Barometric Pressure can be removed by monitoring barometric pressure (eg via a weather station or barometric sensor) and then post processing the absolute water level readings.

The lowest, standard range, absolute pressure sensor offered is 20m, which is suitable for measuring water levels of up to approximately 10m. Absolute sensors will read zero in a perfect vacuum and around 10m in air depending on the atmospheric pressure.

Gauge sensors are suitable for most monitoring applications where water level readings are required. Absolute sensors are suitable for applications where a vented cable is not desirable (eg. Battery pack only sensors).

### 2.5.2 Closed Venting System (CVS)

When pressure sensors are deployed, there can be a difference between the atmospheric temperature and the temperature of the sensor at depth.

This temperature differential causes a pumping effect to occur whereby moist air from the surface is drawn into the sensor through the vent line. This moisture can condense on sensitive electronic components due to warm surface air cooling inside the sensor.



7CVS-001 Closed Vent System

Sealing the system against exposure to the atmosphere and conditioning the existing air in the vent tube can alleviate this problem. Silica desiccant crystals easily absorb moisture thereby drying the air and are used in the closed loop venting system **7CVS-001**.

For all gauge (vented sensors) a Closed Vent System must also be fitted (pictured left). A single 7CVS-001 is designed to handle sensor cable lengths up to **70 metres**. Multiple units may be joined together for greater capacity. Please refer to the Engineering Note in the appendix section on the manual for detailed instructions on the installation of the 7CVS-001. Dimensions (including filter): length x width x height 16cm x 7cm x 5cm.



### **2.5.3 Protective Nose Cones**

A protective copper nose cone (*Greenspan Part # 492-0241*) can be fitted to the pressure transducer to inhibit biological or marine growth on the sensor face. Similarly Greenspan also offer a stainless steel nose cone (pictured left, *Greenspan Part # 492-0246*)



### **2.5.4 BSP Fittings**

Brass BSP threaded adaptors (*Greenspan Part # 492-0238*) can also be fitted to the PS3100 for connection when monitoring pressure in process applications. (Such applications may include, pipeline monitoring, gas bubblers and tanks).

### **2.5.5 Cable Options (Inc Detachable Cables)**

A standard sensor is supplied with a fixed moulded cable entry and bare wire connection. An optional detachable cable is available. The detachable cables have a 7 pin Hirschman connector at the end of the cable opposite to the sensor.

Detachable cables are available in a range of standard lengths and are interchangeable amongst the range of Greenspan sensors. This option can provide benefits and cost savings.

Please refer to the [Engineering Note](#) in the appendix section of the manual for detailed instructions on connecting and disconnecting the detachable cable.

### **2.5.6 Sensor Body**


Sensor is available with a black Acetal or passivated 316 Stainless Steel body. For applications in harsh environments it is recommended that the Acetal body material be specified.

### **2.5.7 Communication Cables**

A communication cable is required to connect the sensor to a PC. The standard cable (5CC-700) has a 7 socket Hirschman connector on one end and a DB 9 on the other end.

## 2.5.8 Optional Serial Output – SDI Adapter Unit

The PS3100 provides on board data Logging of all data and serial output via RS232 to a Laptop or PC using the supplied *SmartCom* software. A feature of the sensor is the ability to also provide serial output in SDI12 format using a small SDI Adapter unit connected to the end of the sensor cable. The SDI12 Adapter unit (Part No 7SDI-1000) provides a standard 3 wire SDI12 output for connection to a third party Data Logger or Process Controller. The PS3100 Sensor can simultaneously provide on-board data logging, as well as act as a standard SDI12 sensor.



When data is requested via SDI-12 the sensor will wake up and take a new set of readings for all channels that are enabled in the sensor. The sensor will then go into a low power, sleep mode. The user can enable or disable channels using SmartCom.

The easiest way to confirm which channels are enabled and what order the data will be returned via SDI-12 is to view the SmartCom monitor screen. The data returned via SDI-12 will be the same channels and in the same order as what is displayed in the SmartCom monitor screen. Note - Memory Used and Memory Used % is not returned via SDI-12.

## 2.5.9 On Board Battery Housing

The Greenspan range of logging Sensors including the PS3100 Sensor may be factory fitted with a non-rechargeable long-life battery pack. This enables the sensor to be deployed at remote sites completely independent of above surface power supplies (no cable connection) and allows for discreet applications. It also functions as a backup power supply in the event of a surface disturbance to the main supply. *Note this option is available for absolute PS3100's only.*


The unit is designed to allow easy access to the battery compartments for battery changeover and is housed in a cylindrical body of approximately the same dimensions as the sensor housing, thus doubling the length of the sensor.



## 2.6 Battery Replacement

Batteries are removed by unscrewing the housing cover, as indicated below and gently levering the battery cells until they slip out. Replacement batteries must be inserted correctly or damage to batteries may occur. Align the +ve on the battery cells with the red indicator on the housing and push batteries in. Batteries are subject to leakage after depletion. The leakage is Thionyl Chloride, a toxic, corrosive non-flammable liquid that can cause damage to equipment and personal injury if in contact with the skin or eyes. Please replace batteries when depleted.

When installing replacement batteries within the battery housing it is necessary to push firmly on the cover until it clicks home over the O rings, after this it can be easily tightened on the thread by hand.



### 2.6.1 Important Battery Information

The type of battery used in the battery pack is Li/MnO<sub>2</sub>, Lithium Thionyl Chloride 3.6V AA cells. A total of nine batteries are required for each sensor battery pack. This configuration supplies a maximum 10.8 volts at 5.2A/Hr and a useful field life, depending on sensor type and logging frequency, of up to 12 months.

#### **REPLACEMENT BATTERIES ARE AVAILABLE FROM GREENSPAN ANALYTICAL**

**Standard AA lithium batteries (Duracell or Energiser Type) are NOT suitable for use in the sensor.**

### 2.6.2 Battery Warnings:

- Do not dispose of batteries in fire, dispose of in appropriate manner.
- Do not short circuit
- Do not expose to water
- Do not crush or puncture
- Do not charge
- Do not over-discharge

To maintain the maximum possible life of the cells before replacement it is strongly recommended that an external power supply is connected to the sensor when downloading data. The power drawn when downloading is at its greatest level, therefore battery depletion will be much more rapid.

Battery life will depend on the battery type as well as the frequency of logging. Connection to a computer will drain the battery supply more quickly due to the higher current imposed by the RS232 serial data communications and will considerably reduce battery life. An additional internal lithium battery maintains logger data at all times but does not sustain the logging state. This battery is not user accessible and will maintain data for up to 10 years.

If the Sensor is fitted with On Board Internal battery pack option and is to be placed in storage it is recommended that the logger be powered down and lithium batteries in the battery pack be removed. To turn off the logger after exiting from SmartCom, disconnect the communications cable and unscrew the battery cover. This exposes the battery compartment to allow removal of the batteries. Removing power will not affect any data remaining in storage so sensors could be downloaded away from the site if required.

## **2.7 Sensor Factory Calibration**


- The transducer is mated to the electronic circuit boards ready for calibration.
- The PS sensor is placed in an environmental chamber and subjected to a matrix of temperature and EC inputs. A typical calibration collects data for each individual sensor at 6 pressure inputs (0, 20, 40, 60, 80, and 100 %FS) at 6 temperatures (0, 10, 20, 30, 40 and 50 deg C).
- A unique calibration curve, for the transducer and electronics set, is calculated from this data and loaded into the sensor.
- The sensor is then re-run through the environmental chamber to ensure that the calibration curve is correct over the entire working range and the sensor meets specifications.
- An extensive range of final calibration and inspection tests are carried out on every sensor.
- The sensor is visually inspected and packed ready for despatch.
- The complete calibration records, sensor history and batch number are placed on file and archived.

## 3 Sensor Operation

### 3.1 Wiring & Connections


The PS3100 is designed as a self-contained data logging sensor. It is normally powered by a 12-24V DC power supply – which can be battery, solar or Mains Plug Pack. The Logger has a large memory capacity and can operate for long periods between field visits. Typically the logged data is collected via laptop using the supplied SmartCom Software.

The following diagram illustrates the typical wiring arrangement for the PS3100.



The SmartCom software provides for configuration of the data logger, collection of logged data, and all other functions. The Greenspan Aquagraph software provides graphical and tabular presentation of logged data and export to spreadsheet format. A brief description of the key components of the SmartCom software is included in this manual. A fully documented online help function is also included within the SmartCom software. It can be started by Hitting F1 key.

Alternatively the PS3100 Sensor can be used as an SDI12 Sensor using the optional Greenspan SDI12 Adaptor.



**Typically the sensor will be connected to a Data Logger or Process Controller which will provide the power and ground connections and provide connections for serial SDI12 output.**

**The Power requirements of the sensor are detailed in the Specifications Brochure.**

The sensor can support internal Data logging and operate as an SDI12 sensor concurrently.

This provides a benefit of having redundant Data Logging capability (internal) while also having the advantage of serial output to a third party Data Logger or Controller that may be connected via some type of telemetry or communications system.

**Please refer to the 7SDI-1000 User Manual for further instructions on how to use the SDI-12 Adaptor with the sensor.**

## 3.2 Software Setup

Communication with the Greenspan logging Sensors is performed through the PC's RS232 serial port using the supplied software: **SmartCom for Windows**. The software supports older style 9 Pin Serial Ports (Com1 & Com2) as well as newer style USB serial Ports. If using a USB port – a USB to Serial 9 pin adaptor must also be used.

**To Load the Software, Place Software CD in drive, and follow all prompts.**

SmartCom for Windows includes two main components:

### 3.2.1 Direct Link

This program is a simpler version of SmartCom which bypasses the Location Explorer Interface to provide connection to a sensor. It is suitable for Direct Com Port Connections only. To connect to a sensor, select the Com Port and click the connect button. Once connected to the sensor the user has the full range of SmartCom functionality available to them but no local Database records of changed or existing Sensor Properties are stored by the software. This version of the Software is recommended for the casual field visit or for technical staff examining the sensor.

### 3.2.2 SmartCom

This is the full version of SmartCom which includes the **Location Explorer** Interface to provide management of individual and/or multiple sensors. This version of the software is the recommended version for customers wanting management of sensor(s) and their logged data.

The following software is required for running the PS3100 Sensor:

1. **SMARTCOM** for Windows runs on a PC, and facilitates manipulation of system setup information, processing and data retrieval. The program is supplied on a CD disk along with support files. It is compatible with WIN98, WIN2000, NT and XP operating systems. Mouse and keyboard operation is fully supported.
2. **Internal firmware**. The logger has a resident program that manages communications, data logging and data retrieval. It is not normally accessible by the user; however it may be upgraded in the field if required.
3. A utility program called **AQUAGRAPH** is also provided on the installation CD. This allows the user to view, graph and export the data. AQUAGRAPH may be accessed from within SMARTCOM. Online help is available from within the program.

The Software will load programs and other files into the following folders:

*C:/Program Files\SmartCom for windows\*

*Several programs will be loaded*

*SmartCom*

*Smartpoll*

*Smart Standby*


*Another folder will also be created*

*C:\Program Files\SmartCom for windows\AquaGraph\*

*With the Graphical display program*

*Aquagraph*


**Note: Administrator Access is required for the Loading of SmartCom Program**




It is recommended that shortcuts be created for the 2 main programs SmartCom & Aquagraph and moved into a dedicated folder on the desktop.

**To Run the Program double click on the SmartCom icon.**

To use the full functions of the software a fully operational sensor with power supply and all communications leads should be available. Either internal or external battery can be used to power the sensor.



**Fully Documented Help Function is available for SmartCom by hitting F1 key while software is running.**




The Main Screen is divided into 2 Panels.

The left side shows the Groups. The right side shows the locations within each Group. (The concept allows all sensor sites to be put together into logical groups – eg by catchment, area, client, sensor type etc)

Locations refer to individual sensors that may be located locally or remotely.

Groups and Locations should be named to reflect the actual station details.




The 4 buttons at the top allow the creation and removal of Groups, and Locations.

To create a new Group – Hit the New Group Button and fill in the preferred name.

To create a new location – highlight the Group then hit the New Location Button.

A wizard will start up to guide you through the process:




The Wizard Has 3 Steps:

Page 1 – requests how to connect to Sensor:  
E.g. Direct Connection via Com port  
(Serial cable between Sensor and PC)

Or

Modem connection via GSM / 3G / landline  
Modem. You will need to enter the phone number  
of the remote modem.

Hit Next Button.



*Step 2 requests Site Details:*

*Location Name  
Instrument Type  
Model Type*



*E.g. this screen will add a new site  
With an PS3100 Logging Sensor  
Named: Black Stump Pressure*

*Hit Next Button.*

*Step 3 Gives a Summary of the entered  
Data and asks for confirmation.*

*Change Details or*

*Hit Finish Button*




*The screen then reverts to the original  
start screen and should now show the  
new Group and the new location.*

*To connect to the Sensor  
Select the Group  
Highlight the location and Hit Connect  
Or  
Double click on the location*

*All information relating to the Site is held  
in files created in the SmartCom folder  
directory.*

**C:\Program Files\SmartCom for  
windows\...  
...\Group Name\Location Name.**



A Pop up will show the steps involved in communicating with the Sensor

- Attempting Wake Up
- Attempting to Connect
- Reading Details
- etc

This will take approx 30 seconds.

If connection cannot be established – check all power and communication connections.


When connection is established the Monitor Screen will appear...

This screen updates every 10 seconds and shows the real time data.

Clock times, Battery Volts, Logging Schedules

Memory Usage should be noted. The Green or Red Leds indicate if Sensor is actively Logging or Not.


There are 2 tabs at the bottom. Hit the Logger Control Tab for all configuration and Logger set up details.




The Logger Control Screen provides access to all configuration, Data retrieval, Logger management and Calibration routines.

The Logger is very flexible and care must be taken to check that Data Logging has been initiated and is working properly, prior to deploying the sensor into a remote site. It is recommended that thorough testing of the sensor be performed in the office to provide confidence and experience with the various logger functions.

All Sensors are supplied from the factory fully calibrated and tested. It is not normally required to perform any additional calibration functions on the sensor.



### 3.2.3 Properties




Key functionality of each of the Property Buttons will be described below.

This front screen provides a summary of all the key sensor information including Serial Numbers, Model Types Firmware Versions etc.

The Channels that are configured for the Sensor.

The Schedules control which channels are logged and the frequency of Logging.

### 3.2.4 Channels




Controls and displays the various Channels that will be displayed on the Monitor Screen.

Channels that are not being used or not required to be displayed can be unchecked and will not be displayed on the Monitor Screen.

(This screen below shows that Internal Battery (Battery Pack) is not used and will not be displayed on the Monitor Screen)

### 3.2.5 External Battery: (Field Supply)



The Sensor will be powered by some type of DC (12-24V) power supply which may be either an internal battery pack or an external battery system. The user must click and select the correct battery type to ensure accurate battery life & voltage calculations.

Select the Option External Battery if using an external 12V battery.

### 3.2.6 Internal Battery: (Battery Pack)

Greenspan Sensors can be configured with an on-board lithium battery pack to provide a complete stand-alone logging sensor.

If the sensor is fitted with the internal on-board battery pack select the Option at left: Battery Pack & Internal Battery.

All batteries whether internal or external (sealed or acid type) will gradually degrade over time The Shelf life of batteries is generally 1-5 years, but can be affected by extreme temperatures.

A feature of the logger allows the user to enter the date of installation and expected Date of Renewal of the batteries, to ensure they are replaced before their due date.

### 3.2.7 Battery Life & Memory Calculator

Provides an estimate of the Memory life and Battery life (days) based on the amount of logging that has been specified via the schedules.

The calculator uses information on the sensor power usage, the frequency of data logging, and the battery capacity to determine how long it will operate before the memory fills, or the batteries go flat.

If using a wrap memory function – the logger will fill with data and commence to overwrite the earliest data.

If using a solar power supply – the battery will likely last until its shelf life determines that it be replaced.

The user can enter the size of the Memory (bytes) & Battery (Ah), and add other power usage such as a Data Modem.

Hit the refresh button to generate new estimates – after changing any of the other parameters. A text box also pops up to provide key information to the user to advise of battery and memory information.

### 3.2.8 Temperature

User configurable items include:

- Unit Name
- Unit Type
- Precision  
(No of decimals)

Data Variation can be ticked, and a value entered, which allows event based logging – should the amount vary by more than the variation between Scan periods.

### 3.2.9 Pressure

User configurable items include:

- Unit Name
- Unit Type
- Precision  
(No of decimals)

Data Variation can be ticked, and a value entered, which allows event based logging – should the amount vary by more than the variation between Scan periods.

Note – All Channels are hard coded into the PS3100 Sensor – but can be individually configured to be included or excluded in Logging Schedules.

### 3.2.10 Schedules

| ScheduleName | Start Method | Stop Method        | Scans Completed |
|--------------|--------------|--------------------|-----------------|
| Main         | 1/01/2002    | Def. STOP DateTime | 0               |
| hourly       | 1/01/2002    | Def. STOP DateTime | 0               |

Schedules are user configured:

- Added
- Removed
- Modified
- Enable
- Disabled

Up to 4 Schedules can be running together.

**Example Schedule – Battery Volts (Field supply)**

Schedule Enabled

Schedule Name  
Field supply

Channels Used in Schedule  
0. Field Suppl

To add channels to a schedule, drag and drop channels from the Tree View, into the adjacent list box max of 8. To Delete channels, right click on channel and select remove.

**Logging Settings**      Averaging

Settings

Scan Time      3 Hours

Record Time      3 Hours

Maximum Number of Scans      0000000

Stop Method

Start Method

*This Schedule shows that Field Supply (External Battery) volts will be measured (Scan Time) every 3 Hours.  
The Data will be logged (Record Time) every 3 Hours.*

*The Schedule is Enabled (box checked) and will operate whenever the Logger is turned on Logging.*

*The start and end time were set to Manual which means that this Schedule will start and stop with the logger.*

**Example Schedule – Water Quality Parameters**

Schedule Enabled

Schedule Name  
hourly

Channels Used in Schedule  
4. Temperature  
5. Turbidity

To add channels to a schedule, drag and drop channels from the Tree View, into the adjacent list box max of 8. To Delete channels, right click on channel and select remove.

**Logging Settings**      Averaging

Settings

Scan Time      1 Hour

Record Time      1 Hour

Maximum Number of Scans      0000000

Stop Method

Start Method


*The Schedule controls the measurement and logging of data as follows:*

*When the logger wakes up to take measurements. (Scan Time)  
When the logger wakes up to Log Data (Record Time)  
Maximum No of Scans (Optional – not recommended)  
Which Channels are included into the Schedule  
(min One, Maximum – all channels)*


*This example shows a Schedule that will wake up and measure Temp, and Pressure – every 1 Hour.  
It will also log the data for these channels every Hour.*

*Note – The Schedule Must be enabled (check box) for it to operate.  
It is possible to have several Schedules in the list that are not enabled and hence will not run when logger is started.*


### 3.2.11 Add New Schedule (Wizard)




*A wizard guides the user through the Add New Schedule procedure.*



*Give the New Schedule a Meaningful Name! (Eg Hourly)*



*Highlight and Drag over the Channels that are required in the new schedule. (This example shows Field supply, Temperature and Pressure Channels)*



*Select the Scan and Record times from drop down list.*

*Sensor will wake up and take a measurement every Scan time, and check if the Data Variation is exceeded for the various Channels. If Data Variation has occurred – Data will be logged.*

*The unit will force a Data Log every Record time.*

*A limit can be placed on the Maximum scans. (Schedule will stop after Max scans reached)*

*Averaging can be configured for the Schedule.*

Select how you would like to start this schedule

Manual Start

Time

< Back      Next >

Start and Stop times can be configured for the Schedule.

It is strongly recommended to use Manual start and stop (default) to ensure that all schedules are operational when logger is started.

Select how you would like the schedule to stop.

Manual Stop

Time

< Back      Next >

### 3.2.12 Add New Schedule – Final Screen

Summary of Schedule

Schedule Name:       Channels Selected:

Scan Time:

Record Time:

---

Start Method:

---

Stop Method:

---

Averaging:       Number of Readings:

Interval between Readings:

< Back      Next >      Cancel      Finish

The final Screen shows summary of the new Schedule.

The Schedule will now appear in the list of Schedules. One or more Channels must be added to the Schedule and it must be enabled to allow it to run.

This example shows that the Schedule named Hourly will measure and log Field supply, Temperature and Pressure Data every hour, whenever the Schedule is enabled, and the logger is turned on. (Logging Active, Green LED on Monitor screen flashing).

### 3.2.13 Averaging

Schedule Enabled

Schedule Name: Hourly

Channels Used in Schedule

- 1, Battery
- 4, Temperature
- 5, Pressure

To add channels to a schedule, drag and drop channels from the Tree View, into the adjacent list box max of 8. To Delete channels, right click on channel and select remove.

Logging Settings

**Averaging**

Averaging Enabled

Number of Readings to Average: 010

Interval Between Readings (secs): 003

Warning : (Number of Readings) x (Interval) <= 3598

Apply Ok Cancel

Averaging enables several readings to be taken and averaged, each time the sensor wakes up at the Scan Period. This allows data fluctuations to be smoothed. For example Logging of Tidal Water levels is typically averaged over say 10 readings say 3 seconds apart to remove the effects of Wind and Swell.

As averaging can increase the time required to generate a reading, a warning will advise the limits of the averaging.

**Note:** Averaging does require the sensor to turn on more frequently at each Scan period and does have an impact on power usage and battery life.

Averaging can be configured for any or all channels.

### 3.2.14 Data Variation (Event Based Logging)

Channel Name: Pressure Unit Name: Metres

Channel Precision: 3 Unit Type: Metres

Data Variation Value Enabled  Data Variation Value: +0000.025

Last Reading: N/A  
Last Logged Reading: N/A

**Advanced**

- Depth: Water Level as the measured by the length of the column of water above the pressure sensor head

To provide Event Based logging the Sensor can be configured to Scan (take a measurement) on a frequent basis, but only Record Data (Log data) if a specific event size (Data Variation) has occurred since the last logged data point.


The Schedule must be enabled.  
The Channel must be in the Schedule and configured with a valid Data Variation.

(This example shows EC will be logged if the reading varies by more than 100EC from the

previous logged reading)

Note if the Data Variation is set to a very small amount, then it is more likely that the data variation will be exceeded, and data will be logged each Scan Time. Hence it is necessary to set the Data Variation based on realistic variations of the Parameter being monitored.

### 3.2.15 User Calibration Screens




Allows re calibration of channels within the logger.

Please consult your sales or technical support specialist before recalibrating the sensor.


The Option "Restore Factory default" allows the sensor to be reset to the original factory calibration.

### 3.2.16 Setting a Datum



The User calibration can be used to set up a sensor so that it reports measurements referenced to a known datum. Once the sensor is installed choose the 1 Point Offset calibration type. When prompted, enter the value that you want the sensor to read.

### 3.2.17 Clear Memory Tab



*This control is used to clear the contents of the Data Logger Memory.*

*Once the memory is cleared – it is gone forever!!!*

*Only perform this function if you have collected and checked all the data from the sensor.*


*Aquagraph can be used to view and check data before deleting the memory contents.*

If using the memory Wrap function – the memory will fill and then overwrite the oldest (earliest data)

***If not using the Memory Wrap function – the memory will fill and then no more data will be logged. (i.e. all the data from the start of the logging period will be retained until it is cleared)***

### 3.2.18 Set Clock

The logger has an internal clock to provide time stamping for all data logging. The clock can be reset to the correct local time via two different methods.




### 3.2.19 PC Clock

Option 1 – allows the Logger Clock to be set to the clock time of the PC being used.

PC Clock shown in red (left)  
Logger Clock shown in green (right)

The Logger clock is displayed in the top of the screen – and should be checked to confirm it is correct.



### 3.2.20 Manual Entry Clock

Option 2 – allows the Logger Clock to be set to a User Manually Entered time.


Screen shows the drop down menus for the user to enter the correct local time and date settings into the Logger Clock.

Logger Clock shown in green (right)

The Logger clock is displayed in the top of the screen – and should be checked to confirm it is correct.



### 3.2.22 Collect Data




*This allows collection of the logged data. Various options allow collection of all Data, New Data or Data between specific dates. Data can be downloaded to new files, or appended to existing files.*

*Data is not deleted from the Logger during this process and can be re-collected many times.*

*The Data Logger does not need to be stopped for the collection process, and normal logging of data will continue while the data downloads.*

*The Data is extracted in Binary Format and requires the Aquagraph program to decode into text format.*



*Data collection can take up to several minutes for direct connection (serial cable) or longer for remote modem connections.*

*It is always more efficient to perform regular data collections of New Data – and append to a Master File, rather than try to collect very large parcels of data.*

*Once complete the Data can be viewed in Tabular and Graphical format by hitting the View Data in Aquagraph button.*

### 3.2.23 Aquagraph – Tabular View


| Date      | Time       | Elapsed Time (mins) | Schedule | Parameters          |              |               |         |               |
|-----------|------------|---------------------|----------|---------------------|--------------|---------------|---------|---------------|
|           |            |                     |          | Temperature Celsius | EC RAW uS/cm | EC NORM uS/cm | TDS TDS | Salinity mg/l |
| 7/02/2008 | 1:40:00 PM | 0.0000              | 1        | 22.12               | -5046        | -5304         | 15      |               |
| 7/02/2008 | 1:45:00 PM | 5.0000              | 1        | 22.13               | -4339        | -4559         | 15      |               |
| 7/02/2008 | 1:50:00 PM | 10.0000             | 1        | 22.13               | -5761        | -6054         | 15      |               |
| 7/02/2008 | 1:55:00 PM | 15.0000             | 1        | 22.14               | -4340        | -4560         | 15      |               |
| 7/02/2008 | 2:00:00 PM | 20.0000             | 1        | 22.14               | -4340        | -4560         | 15      |               |
| 7/02/2008 | 2:05:00 PM | 25.0000             | 1        | 22.14               | -3629        | -3812         | 15      |               |
| 7/02/2008 | 2:10:00 PM | 30.0000             | 1        | 22.13               | -3626        | -3811         | 15      |               |
| 7/02/2008 | 2:15:00 PM | 35.0000             | 1        | 22.14               | -4339        | -4559         | 15      |               |
| 7/02/2008 | 2:20:00 PM | 40.0000             | 1        | 22.21               | -3642        | -3822         | 15      |               |
| 7/02/2008 | 2:25:00 PM | 45.0000             | 1        | 22.22               | -3644        | -3823         | 15      |               |
| 7/02/2008 | 2:30:00 PM | 50.0000             | 1        | 22.22               | -2934        | -3077         | 15      |               |
| 7/02/2008 | 2:35:00 PM | 55.0000             | 1        | 22.21               | -4355        | -4569         | 15      |               |
| 7/02/2008 | 2:40:00 PM | 60.0000             | 1        | 22.21               | -3643        | -3822         | 15      |               |
| 7/02/2008 | 2:45:00 PM | 65.0000             | 1        | 22.21               | -5066        | -5315         | 15      |               |

Shows all logged data from the sensor.

Data can be exported to excel spreadsheet.

To generate graph – highlight column(s) of interest and hit graph button.

### 3.2.24 Aquagraph – Graphical View




Graphs will self scale across the dates and ranges requested.

Multiple channels can be plotted on single graph, as well as stacked graphs for many parameters.

Various tools are available from Menu for highlighting points, adding text, zoom function etc.

Graph can be exported as a graphic (.bmp) file for reports etc.

### 3.2.25 Aquagraph – Tools & Options



Battery Details can be included.

Rainfall can be included in various formats  
 Instant readings (each tip)  
 Cumulative Total  
 Daily Rainfall total

Other options support display of flow meter readings, and conversion of pressure/level to Depth of Water above or below specified Datum levels.

Fully documented Help Function is available for Aquagraph by hitting F1 key while software running.

## **3.3 Configuration & Testing**

The Factory supplied Calibration Certificate will provide detailed information on the calibration of the sensor and should be retained for future reference.

To maintain high quality control over monitoring programs, it is recommended that calibration is checked every 3-6 months. If re-calibration is required a method is presented here using Standard Solutions. Alternatively sensors may be returned to an authorised Greenspan agent for re-calibration.

### **3.3.1 Quick Check and Re-Calibration Method**

The sensor can be checked by use of known Pressure inputs, usually from an accurate digital pressure calibrator or dead weight tester. However a quick check can be done at any time by reading the sensor in air.

### **3.3.2 Quick check for a gauge sensor**


1. Remove the sensor from the water, ensure the sensor is clean and dry and the vent tube at the top of the cable is not sealed.
2. Connect the sensor to a PC and run SmartCom.
3. Once connected the monitor screen will display the readings. The Pressure value should read Zero +/- 0.1% of the full scale range of the sensor.
4. This confirms that the sensor electronics has remained stable and no further action should be required if the sensor is within +/- 0.1 % FS.

### **3.3.3 Quick check for an absolute sensor**


1. Remove the sensor from the water, ensure the sensor is clean and dry.
2. Connect the sensor to a PC and run SmartCom.
3. Obtain the barometric pressure for the location \* where the check is being performed.
4. Once connected the monitor screen will display the readings. The Pressure value should read the barometric pressure +/- 0.1% of the full scale range of the sensor. (1hPa = 0.010216 m H<sub>2</sub>O @ 20°C)
5. This confirms that the sensor electronics has remained stable and no further action should be required if the sensor is within +/- 0.1 % FS.

\* The accuracy and calibration of the barometer should be considered when making the comparison.


### 3.3.4 Re-Calibration Method Using Digital Pressure Calibrator or Dead Weight Tester




1. Ensure the sensor is clean and dry.
2. Connect the sensor to the calibrator using a BSP fitting (See section 2.5.4).
3. Provide power to the sensor, connect sensor to a PC with appropriate communication cable.
4. Run SmartCom for Windows.
5. In SmartCom for Windows, select User Cal from Logger Control menu.
6. Select the PressureCal Channel.
7. Select 2 point Span and Offset Calibration Type.




8. An advice box will appear on the screen, click yes.



9. Apply the low pressure input to the sensor from the calibrator. Typically this will be zero for a gauge sensor.
10. The screen should display a window to allow entry of the new low value, type in the new value to be read by the Sensor for zero, eg: (0000.00), click OK.



11. Apply the high pressure input to the sensor from the calibrator. Typically this will be full scale for a gauge sensor.
12. The screen should display a window to allow entry of the new high value, type in the new value to be read by the Sensor for the high value, eg: (0005.00), If the calibrator cannot set an output in m of water a conversion will need to be calculated and entered, click OK.



13. An advice screen will appear, click ok

Note – Absolute pressures must be set and entered when calibrating an absolute sensor.

## **4 Field Deployment Considerations**

For applications in harsh environments it is recommended that the optional Acetal casing be specified.

The sensor head should be periodically inspected for fouling, and can be cleaned with fresh water and damp cloth. In marine environments crustaceans may need removal at regular intervals.

The body should always be fully immersed under the water to ensure the electronic module is at water temperature and to avoid any possible anodic/cathodic action taking place on the stainless body due to the oxygen difference across the boundary.

Care should be taken if clamps are to be attached to the Stainless Steel body as the depletion of oxygen to the clamp/probe interface can cause corrosion due to anodic/cathodic action. It is recommended that Acetal body sensors be used if clamping.

Sensors should generally be installed such that they can be easily and safely removed for cleaning, servicing. For environmental applications the sensor can often be mounted inside a section of PVC or steel pipe which enters the water body. The sensor can then be slid down inside the pipe until the sensor head just protrudes into the water body. This provides a high degree of protection for the sensor from environmental (sunlight, heat, flood debris etc) as well as from other influences such as Cattle, vandalism etc. Most sediment transport occurs during storm events and flood conditions. Protection from floating debris damage is an important consideration along with adequate tethering of sensors.

### **4.1 Cabling Considerations**

Care should be taken with installation and field servicing to ensure the cable is not subjected to persistent pulling snagging or severe compression. Cyclic loading of the cable should also be avoided through careful sensor deployment. Additional stilling wells or mounting brackets may be required to prevent sensor movement which may cause long term cable movement. Where cable runs are required which may be subject to environmental effects (heat, water movement, sunlight, flood debris etc) it is advisable to protect the sensor cable inside a slightly larger diameter conduit such as PVC, steel or polyethylene. This also allows the sensor cable to be pulled out – should a sensor change-over be required at the site. Maximum cable runs up to several hundred meters are possible without affecting electrical signals.

## **4.2 Typical Sensor Installations**

1. Edge of river/stream/lake embankment.
2. Mounted within a stilling well off stream from main flow.
3. Mounted within drainage channels/pipes.
4. Suspended from dam walls.
5. Sensor anchored to bed of lake/stream.

## **4.3 Field Installation must ensure:**

- The sensor is anchored or held in position or located so it is not subject to any movement during normal operations.
- Sensor is protected from direct sunlight to avoid high temperature fluctuations
- Sensor is protected against high turbulence and possible debris loading during flow events

## **4.4 Other Considerations**

Environmental compatibility should be checked before using the sensors and advice sought from Greenspan if any doubt exists. The sensor utilises some 316 stainless components that are suitable in a majority of situations but care should be taken against possible corrosion in high Chloride, Sulphate or Ferric solutions. The body should always be totally immersed under the water to ensure that the sensor is at water temperature and to also avoid any possible anodic/cathodic action taking place on the components at the water-air interface. If using clamps to mount the sensor – these should be of a type that evenly clamps the sensor body without excessive loading that could damage the sensor body.

## **4.5 Guidelines for cleaning equipment**

The sensor may be cleaned using a soft cloth, mild detergents and warm water. If the sensor shows signs of marine growth a light biocide can be used to clean and kill any biological growth on the sensor.

## 5 Appendix A -Additional Information

### 5.1 Specifications

|   |  |
|---|--|
| Measurement Technique   | Pressure ½ " Ceramic Capacitance Sensor<br>Temp – integrated precision thermistor                                  |
| Operating Ranges  | Temp: 0 – 50°C<br>Pressure: Max Over Range – 4 times full scale range  |
| Standard Pressure Ranges available                                  | Gauge: 2.5, 5, 10, 20, 40, 75 100m<br>Absolute: 20, 40, 75, 100m   |
| Other Ranges Available  | Yes – Calibration charge may apply – refer sales office  |
| Sensor Outputs  | Internal Data Logger – serial data via Smartcom software<br>Optional adaptor provides SDI12 serial output (3 wire) |
| Overall Accuracy (combined linearity, hysteresis and repeatability) | Pressure: +/- 0.1% full scale range<br>Temp: +/- 0.2°C   |
| Long Term Stability   | 0.2% full scale per annum  |
| Resolution  | Depth: 0.001m, Temp: 0-50°C  |
| Storage Temperature   | -5°C -+60°C  |
| Cable Type  | Polyurethane sheathed cable, OD 8mm with 3mm vent tube, moulded entry, HS7 connector for serial connection         |
| Cable Lengths   | 10, 20, 30, 50, 100, 150m (32, 65, 100, 165, 325, 490 FT)  |
| Power Supply  | 8-30Vdc (at sensor), or On-board battery pack (optional)   |
| Power ESD Protection  | 2000 volts   |
| Current Consumption   | Sleep<0.2mA, logging 20mA, and communicating 30mA.   |
| Sensor warm up time   | 5 seconds  |
| On-board battery pack (option)                                      | Housing screws to sensor size (OD x L) 47mm x 250mm  |
| Battery capacity  | 9 x Lithium AA (3.6Volt) – Total Capacity 5.2Ah @ 10.8v  |
| Typical Field Life  | Over 12 months remote operation @ 15 minute logging  |
| Internal data logger  | Non-volatile, battery-backed RAM with real-time clock  |
| Memory size   | 2Mb capacity, with user selectable wrap function   |
| Measuring Units   | User definable (metric and Imperial US units)  |
| Data Storage  | 250,000 Readings (Typically 5 minute data for >12 months)  |
| Logging frequency   | User selectable from 1 second up to once per day   |
| Dimensions (L x OD)   | 355mm x 47mm (14" x 1.78")   |
| Weight  | 500g plus cable weight (665g per 10m length)   |
| Wetted Materials  | UPVC, acetal, 316 passivated stainless steel, polyurethane, viton  |
| CVS Dimensions: length x width x height (including filter)          | 16cm x 7cm x 5cm   |

## 5.2 Quick Start Guide – PS3100 Pressure Sensor



### PS3100 Pressure Sensor

# CERTIFICATE of CONFORMANCE

**Customer:** "Click here & type Customer name"  
**Model No.** PS3100 Pressure Sensor (Tyco Environmental Systems material # 700-3100)  
**Sales Order Reference:** "Click here & type SO Reference"  
**Serial Number:** "Click here & type S/N"

#### Product Information



PS3100 Pressure Sensor

|                            |  |
|----------------------------|--|
| <b>Range</b>               | "Type Range here" H <sub>2</sub> O     |
| <b>Output</b>              | RS232                                  |
| <b>Pressure Linearity*</b> | +/- 0.05% FS                           |
| <b>Pressure Accuracy</b>   | +/- 0.10% FS<br>(over 0-50 °C)         |
| <b>Temp Accuracy</b>       | +/- 0.2 °C                             |
| <b>Sensor Type</b>         | Gauge/Absolute (delete applicable)     |
| <b>Cable Length</b>        | "Click here & type Cable length" M     |
| <b>Firmware Version</b>    | RC 3.11                                |
| <b>Ext Supply Voltage</b>  | 9 - 30 VDC                             |
| <b>Connection Code</b>     | HS7/Detachable/CX9 (delete applicable) |

\*Combined Linearity, Hysteresis and Repeatability

#### User Notes

1. Australian Standard, AS1376 is used to convert kPa to metres of water.
2. (1kPa = 102.15 mm water) at 20°C
3. Apart from battery access, dismantling of the sensor will void the warranty. Contact your agent for technical advice.
4. The sensor is protected against reverse polarity connection.
5. The sensor is fitted with a lightning protector/surge device.
6. The sensor is compensated for temperature induced errors over the range 0 -50°C.
7. The sensor turn on time is factory set to 5 seconds.

Inspected By: \_\_\_\_\_ 1<sup>st</sup> January 2009

**Manufactured By:**  
 Tyco Environmental Systems (Greenspan Analytical Plant)  
 22 Palmerin Street  
 WARWICK QLD 4370  
 AUSTRALIA Phone: + 61 (0)7 46601888




# Quick Start Guide – PS3100 Pressure Sensor


Procedures for connecting and configuring the Greenspan PS3100 Sensor. Further details in the Sensor Manual included on the Greenspan CD.

### You should have received:

- Certificate of Conformance Sheet
- CD Manuals & Software
- Greenspan PS3100 Sensor



As an option the sensor may be configured with on board battery housing



*PS3100 provides a complete logging pressure sensor. Communication and configuration of the sensor is via the Smartcom Software on the supplied CD.*

PS3100 Quick Start Guide

## 5.3 Engineering Note – Detachable Cables

### ENGINEERING NOTE

RELEASE DATE: 28/4/2008

SUBJECT: Connecting and Disconnecting the Detachable Cable for Greenspan Sensors (packaged in 47mm & 65mm Tubes)

#### IMPORTANT NOTE:

You have been supplied with:

2 x O-Rings (Greenspan part number 011-OR16X2.5)


2 x Grub Screws (Greenspan part number 512-M4X6SS316P)

1 x Allen Key


**Prior to connecting your sensors to the detachable cable, please ensure that the detachable cable entry has been fitted with O-Rings supplied as detailed above. Ensure O-Ring grooves are greased and then slide O-Rings and fit into the recessed grooves.**




Refer to the following instructions for **connecting** to your sensor to the detachable cable



1. The sensor connector is exposed and ready to connect to the cable connector.



2. Align the cable connector with the sensor connector and push together. The connector is polarized and will only fit together one way.



3. Once connectors are completely pushed together, the rocker arm on the sensor connector will clip in with the cable connector and secure the connection.



4. Carefully feed the connection back into the sensor tube taking caution not to crimp or damage any wiring. Ensure the O-Rings have been fitted to the moulded cable entry as directed at the beginning of this document



5. Twist carefully to align the grub screw holes. Firmly push the tube back onto the moulded cable entry



7. Screw in two (2) 512-M4X6SS316P grub screws to secure the cable and the sensor with supplied Allen key



8. The cable is now securely fitted to the sensor.


Refer to the following instructions for **disconnecting** your sensor to the detachable cable




1. Sensor with cable attached and ready for cable disconnection.




2. Remove the two (2) grub screws (Greenspan part #512-M4X6SS316P) using the Allen key supplied. (N.B Allen key is supplied in spare parts kit with original packaging)



3. CAREFULLY pull tube away from cable mould. Once the tube has been pulled away from cable mould, the connector will be revealed as pictured left.



4. Release locking device on the connector by gently pinching rocker arm on the cable connector and pull apart.



5. The sensor and cable are now detached.

**Note:** When the cable is not connected to a sensor, please ensure the supplied vent cap is fitted to the vent tube (if applicable)